

Autoclave on third floor:

Cycle Select: Liquid (for liquids)

Close door securely

Light should go on for 'auto'

'Door locked' light should be on

'Start'

Jacket pressure should rise to 20

then will really start